

Declaration of Conformity

According to ISO/IEC 17050-1:2010

We as a manufacturer:

HK Instruments Oy
Keihästie 7
FIN-40950 MUURAME, Finland
www.hkinstruments.fi

Declare that the products:

DPI
DPI+/-300 (-1R/2R-AZ) -D
DPI+/-500 (-1R/2R-AZ) -D
DPI2500 (-1R/2R-AZ) -D

Conform to the following Directives:

EMC Directive 2014/30/EU
RoHS Directive 2011/65/EU
WEEE Directive: 2012/19/EU
Product Safety: 2001/95/EC
LVD Directive: 2014/35/EU

According to the following relevant standards:

IEC/EN61010-1: 2001 (Second edition) Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements

EN 61326-1: 2006 Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements

EN 61326-2-3: 2006: Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-3: Particular requirements - Test configuration, operational conditions and performance criteria for transducers with integrated or remote signal conditioning

According to the following EMC test methods:

EN 61000-4-2: 2001
EN 61000-4-3: 2002
EN 61000-4-4: 2004
EN 61000-4-5: 2001
EN 61000-4-6: 2001

Supplementary information:

These products comply with the requirements of the Low Voltage Directive 2014/35/EU and the EMC Directive 2014/30/EU, and carries the CE-marking accordingly.

Muurame, June 2016



Tuomo Kalliomäki
CEO